

# Product datasheet

## AAV2 ELISA Control

### Short overview

<b>Cat. No.</b>	PRAAV2-C
<b>Quantity</b>	1 vial (0.8E+09-1.4E+09 capsids)
<b>Concentration</b>	1.7E+09-2.8E+09 capsids/ml after reconstitution in 500 µl ASSB 1x (please find the lot-specific concentration on the CoA and on the vial)

### Product description

<b>Formulation</b>	Lyophilized, empty AAV2 capsids. Reconstitute in 500 µl ASSB 1x (provided with PROGEN's AAV2 Titration ELISA), incubate 5 min at RT and mix by rolling 5 min. Avoid vortexing! The final solution contains stabilizing protein, phenol red and ASSB 1x buffer.
<b>Stability</b>	4 weeks at 2-8°C after reconstitution in ASSB 1x
<b>Storage</b>	2-8°C
<b>Intended use</b>	Research use only
<b>Application</b>	ELISA

### Applications

<b>ELISA</b>	As a positive control in ELISA, a 1:4 dilution in ASSB 1x (provided with PROGEN's AAV2 Titration ELISA) and analysis at least in duplicates is recommended.
--------------	---

### Background

The AAV2 ELISA Control consists of fully assembled, empty AAV2 capsids. The concentration is lot specific. The AAV2 ELISA Control can be used as a positive control with PROGEN's AAV2 Titration ELISA (PRATV & PRAAV2R). If you require different lots of the AAV2 ELISA Control and the Kit Control included in your PROGEN AAV2 Titration ELISA, please enquire to check availability. Reading of the AAV2 ELISA Control from the standard curve (i.e. Kit Control, included in PROGEN's AAV2 Titration ELISA) is influenced by inter- and intra-assay, but also by inter-lab variances. Therefore, it is recommended that each laboratory determines its own acceptable range of recovery. The AAV2 ELISA Control has been calibrated on the ATCC AAV2 reference standard material (RSM, VR-1616). For further information, please see the publication by Lock et al., 2010. PublicationsLock M, McGorray S, et al. Characterization of a recombinant Adeno-Associated Virus Type 2 reference standard material; Hum Gene Therapy 21:1273-1285 (2010).

### Product images



AAV2 ELISA Control

## References

Publication	Species	Application
<a href="#">Lock M, McGorray S, et al. Characterization of a recombinant Adeno-Associated Virus Type 2 reference standard material; Hum Gene Therapy 21:1273-1285 (2010).</a>		
<a href="#">Grimm, D. et al. Titration of AAV-2 particles via a novel capsid ELISA: packaging of genomes can limit production of recombinant AAV-2. Gene Ther. 6, 1322â€“30 (1999).</a>	AAV2	ELISA,dot blot,IP,ICC-IF