

Product datasheet

AAV2 ELISA Control

Short overview

Cat. No.	PRAAV2-C	
Quantity	1 vial (0.8E+09-1.4E+09 capsids)	
Concentration	1.7E+09-2.8E+09 capsids/ml after reconstitution in 500 µl ASSB 1x (please find the lot-specific	
	concentration on the CoA and on the vial)	

Product description

Formulation	Lyophilized, empty AAV2 capsids. Reconstitute in 500 µl ASSB 1x (provided with PROGEN's AAV2 Titration ELISA), incubate 5 min at RT and mix by rolling 5 min. Avoid vortexing! The final solution contains stabilizing protein, phenol red and ASSB 1x buffer.	
Stability	4 weeks at 2-8°C after reconstitution in ASSB 1x	
Storage	2-8°C	
Intended use	Research use only	
Application	ELISA	

Applications

ELISA

As a positive control in ELISA, a 1:4 dilution in ASSB 1x (provided with PROGEN's AAV2 Titration ELISA) and analysis at least in duplicates is recommended.

Background

The AAV2 ELISA Control consists of fully assembled, empty AAV2 capsids. The concentration is lot specific.

The AAV2 ELISA Control can be used as a positive control with PROGEN'S AAV2 Titration ELISA (PRATV & PRAAV2R). If you require different lots of the AAV2 ELISA Control and the Kit Control included in your PROGEN AAV2 Titration ELISA, please enquire to check availability.

Reading of the AAV2 ELISA Control from the standard curve (i.e. Kit Control, included in PROGEN's AAV2 Titration ELISA) is influenced by interand intra-assay, but also by inter-lab variances. Therefore, it is recommended that each laboratory determines its own acceptable range of recovery.

The AAV2 ELISA Control has been calibrated on the ATCC AAV2 reference standard material (RSM, VR-1616). For further information, please see the publication by Lock et al., 2010.

Publications

PROGEN Biotechnik GmbH | Maaßstraße 30 | D-69123 Heidelberg Tel.: +49 (0) 6221 8278-0 | Fax: +49 (0) 6221 8278-24 | Email: info@progen.com | Web: www.progen.com 2024 April 19 / Version: PRAAV2-C/DS-090420ibg | Page 1 Lock M, McGorray S, et al. Characterization of a recombinant Adeno-Associated Virus Type 2 reference standard material; Hum Gene Therapy 21:1273-1285 (2010).

Product images



AAV2 ELISA Control

References

Publication	Species	Application
Lock M, McGorray S, et al. Characterization of a recombinant		
Adeno-Associated Virus Type 2 reference standard material:		
Hum Gene Therapy 21:1273-1285 (2010).		